

**g.tec medical engineering GmbH**  
Sierningstrasse 14  
4521 Schiedlberg, Austria  
Tel. 0043 7251 22240 0  
Internet: www.gtec.at



## Konformitätserklärung Declaration of Conformity

### Product details:

Product name	<b>g.Nautilus Research / g.Nautilus Research Sahara and components</b>
Type	<b>8, 16, 32, 64 -Channel Wireless Amplifier dry and wet systems</b>
Order number	<b>5130, 5132, 5134, 5136, 5137, 5135, 5133, 5131, 5408, 5416, 5432, 5464, 5508, 5516, 5532, 5564</b>

### Classification:

Safety class	<b>II (according to IEC 60601-1)</b>
Type of applied part	<b>BF (according to IEC 60601-1)</b>
Protection against mechanical distortion and liquids	<b>IP44 Headset (according to EN 60529) IP40 Base Station (according to EN 60529)</b>
Operation Mode	<b>S1 (permanent operation, according to IEC 60601-1)</b>

### Used standards:

Harmonized standards	<b>EN 60601-1:2006+A1/2013 EN 60601-2-26:2015 (partly fulfilled) EN 60601-1-2:2007+AC2010 EN 62304:2006</b>
----------------------	---

The manufacturer declares in sole responsibility that the device mentioned above under Product is in conformity with Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment and are conform with the general requirements for safety 2001/95/EC and electromagnetic compatibility 2014/30/EU.

Disclaimer: g.Nautilus Research is not a certified medical device and must not be used as a medical device according to the Medical Device Directive (MDD 93/42/EEC).



**Schiedlberg, 27.09.2019**  
**Validity 1 year**

**Dr. Günter Edlinger,**  
**Chief Executive Officer**